



Langley Research Center Safety Alert

SA-06-93

Date: December 8, 1993

TO: Distribution
FROM: 429/Safety Manager, Head, Risk Management Branch, SSQRD
SUBJECT: Instrument Calibration

All programs and organizations utilizing instrumentation that is on a mandatory calibration sequence should be aware that the NMIS data base for the LaRC Calibration Recall System experienced a failure which automatically removed many instruments from the LaRC Calibration Recall System. As many of these instruments are mandated by the Nuclear Regulatory Commission to be calibrated at specific time intervals, all efforts should be made to insure that your instruments are within calibration. Failure to do so could result in a Notice of Violation from the Nuclear Regulatory Commission. Please contact Mr. Phillip D. Babb, Radiation Safety Officer, at extension 43210 if additional support is required.

V. William Wessel
V. William Wessel
4SAFE

Distribution:
Facility Safety Heads
Facility Coordinators
Contractors Safety Reps.

421/JVC *JVC*